

INFORMATION FOR LIMITED DISTRIBUTION



Technical Bulletin

Ref: **TB-16017-ENG**

Product: INOmax DS_{IR}® Plus MRI

Affected parts: Operation Manuals

Subject: **MR Conditional Ventilator/Breathing Systems Compatibility List**

Issue date: **October 2016**

Priority: Low

Classification: Information for Limited Distribution

The INOmax DS_{IR} Plus MRI has been validated for use with the following ventilators/breathing devices in the 1.5 and 3.0 tesla scanner rooms:

Ventilators/Breathing Systems Validated for use in the 1.5 and 3.0 tesla Scanner Rooms	
Manufacturer	Model
Bio-Med Devices	MVP-10
CareFusion (Formerly Pulmonetic Systems)	LTV 1200 MRI
Hamilton	MR1
Maquet	Servo-i (MR Conditional)
Nasal Cannula	Salter Labs, Infant, REF 1601, with 7 foot supply tube (or equivalent)
Smith's Medical	babyPAC

For technical assistance regarding the INOmax DS_{IR} MRI, please contact technical support at 1-877-566-9466 (North America) or your specific country manager.

Mallinckrodt Manufacturing LLC
6603 Femrite Drive,
Madison, WI 53718-6801 USA

Mallinckrodt, the "M" brand mark and the Mallinckrodt Pharmaceuticals logo are trademark of a Mallinckrodt company. Other brands are trademarks of a Mallinckrodt company or their respective owner. ©2016 Mallinckrodt